K100615

510(k) SUMMARY

MAR 3 0 2010

Sponsor/Submitter:	Arstasis, Inc. 1021 Howard Avenue, Suite C San Carlos, CA 94070
Contact Person:	Su-Mien Chong Acting VP, Regulatory & Clinical Affairs Phone: (650) 704-1632 Fax: (650) 594-4326
Date of Submission:	March 3, 2010
Device Trade Name:	Arstasis^{one} Access System
Common Name:	Catheter Introducer
Device Classification:	Class II
Regulation Number:	21 CFR 870.1340
Classification Name:	Catheter Introducer
Product Code:	DYB
Predicate Device:	Arstasis ITG Vascular Access System (K091006)
Device Description:	Arstasis^{one} is a device that is comprised of a sheath, anchor mechanism, shaft and handle with control features.
Indications for Use:	Arstasis^{one} Access System is intended to provide access for the percutaneous introduction of devices into the peripheral vasculature and to promote hemostasis at the arteriotomy site as an adjunct to manual compression. The Arstasis^{one} Access System is indicated for use in patients undergoing diagnostic femoral artery catheterization procedures using 5F or 6F introducer sheaths.
Technological Characteristics	Arstasis^{one} is designed to create a shallow access path through the arterial wall for the guidewire to enter the vessel lumen.
Summary of Substantial Equivalence:	Bench testing was performed on the subject device as follows: functionality, deployment and release forces, flexibility, torque loading as well as tensile, compressive and torque strengths. Biocompatibility testing was successfully conducted pursuant to ISO-10993-1, Biological Evaluation of Medical Devices Part 1: Evaluation and Testing (1995).

Prior testing included preliminary animal studies (non GLP) and cadaver assessments[1], as well as clinical investigations.[2] Multiple clinical evaluations were conducted. The short term safety and clinical performance of the device were established. The long term safety, as well as the ability to access and re-access, was retrospectively studied in a smaller cohort of patients.

In summary, the cumulative data provided herein demonstrate that the Arstasis^{one} Access System is substantially equivalent to its predicate in providing access to the arterial lumen and facilitating the introduction and placement of devices into the peripheral vasculature and achievement of hemostasis.

[1] The preliminary Animal Studies and Cadaver Assessments were conducted using prototypes of a similar design and configuration.
[2] Clinical investigations were conducted on an earlier device version of similar design and configuration.

DEPARTMENT OF HEALTH & HUMAN SERVICES

Public Health Service

Food and Drug Administration
10903 New Hampshire Avenue
Document Control Room W-O66-0609
Silver Spring, MD 20993-0002

APR 29 2010

Ms. Su-Mien Chong
Acting Vice President, Research and Development
Arstasis, Inc.
1021 Howard Avenue, Suite C
San Carlos, CA 94070

Re: K100615
Trade/Device Name: Arstasis^{one} Access System
Regulation Number: 21 CFR §870.1340
Regulation Name: Catheter, Introducer
Regulatory Class: Class II (two)
Product Code: DYB
Dated: March 3, 2010
Received: March 4, 2010

Dear Ms. Chong:

This letter corrects our substantially equivalent letter of March 30, 2010.

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please go to http://www.fda.gov/AboutFDA/CentersOffices/CDRH/CDRHOffices/ucm115809.htm for the Center for Devices and Radiological Health's (CDRH's) Office of Compliance. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm.

Sincerely yours,



Bram D. Zuckerman, MD
Director
Division of Cardiovascular Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

APPENDIX B: INDICATIONS FOR USE STATEMENT

510(k) Number (if known): K100615

Trade Name: Arstasis[one] Access System

Common Name: Catheter Introducer

Indications For Use: The Arstasis[one] Access System is intended to provide access for the percutaneous introduction of devices into the peripheral vasculature and to promote hemostasis at the arteriotomy site as an adjunct to manual compression. The Arstasis[one] Access System is indicated for use in patients undergoing diagnostic femoral artery catheterization procedures using 5F or 6F introducer sheaths.

Prescription Use ___√___ (Part 21 CFR 801 Subpart D) AND/OR Over-The-Counter Use _____ (21 CFR 801 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE OF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

(Posted November 13, 2003)

Susan R. Vichnes
(Division Sign-Off)
Division of Cardiovascular Devices

510(k) Number K100615